MHolland

Face shields can be made from any clear styrenic, PC or PET. Mask frames can be PS or PP and often use an elastomer SEBS or similar skin interface.

Bed linings are often made with a breathable film to prevent soil staining.

All housings and bezels of the **ventilator** use PC blends and must resist hospital disinfectants.

Masks are often made using PP fiber.

Gowns can be PP fabric or blown film.

Tubes are made of PVC and some parts are TPU.

Bed structural parts are talc and/or glass filled PP, all electronics are PC blends.

Wheel bases are long glass PP and can be short glass nylon.